6

Notice of Allowability	Application No.	Applicant(s)
	09/690,473	HAARALA ET AL.
	Examiner	Art Unit
	Michael J. Hayes	3767
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.		
1. This communication is responsive to <u>BPAI decision of 6/7/05</u> .		
2. The allowed claim(s) is/are 60,61,63-68,70-72,74-79,113 and 114.		
<ul> <li>3.</li></ul>		
6. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.		
Attachment(s)  1. ☐ Notice of References Cited (PTO-892)  2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)  3. ☐ Information Disclosure Statements (PTO-1449 or PTO/SB/O Paper No./Mail Date  4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material	6. ☐ Interview Summary Paper No./Mail Dat 98), 7. ☐ Examiner's Amenda	e nent/Comment ent of Reasons for Allowance Michael J Hayes
		Primary Examiner Art Unit: 3767

## Claims Allowable

Claims 60, 61, 63-68, 70-72, 74-79, 113, and 114 are allowed in view of BPAI decision, paper 9/19/05.

## Reasons for Allowance

Claims 60, 61, 63-68, 70-72, 74-79, 113, and 114 are allowed because the recited combination of elements of an access port to be implanted comprising a body portion having upper and lower attachable body parts with a self-sealing septum between, forming a reservoir, a reservoir outlet, with an entry site on the body portion disposed opposite the outlet and configured to permit insertion of a guidewire or stylet through the body portion and into the outlet where the septum exterior forms a portion of the exterior surface of the device (claim 60) or with an entry site and access site configured to permit access to the reservoir where an outer surface of the septum forms a portion of the exterior of the surface of the device is not found nor fairly taught in the prior art of record. See BPAI decision, paper 9/19/05.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Hayes at (571) 272-4959. The examiner can usually be reached Monday -Thursday, 7:00-4:30, and on alternate Fridays. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Sirmons, can be contacted at (571) 272-4965. The fax number for submitting official papers is (571) 273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

mjh

17 October 2005

Page 2

PRIMARY EXAMINER